

WCM-2000 SOFTWARE CHANGES

This document is intended to keep a record of major modifications made to the WCM-2000 software.

07/14/10

Fixed problem with output port name and protocol for SeaWcm. It didn't allow for the change to be made from the GUI. Also gave invalid port name/protocol.

04/23/10

Added support for NOAA IWG1 data sentences.

04/20/10

Fixed and improved auto start time. Fixed play/stop issue. Fixed expert mode issue.

02/17/10

Change SeaWcm input to be more forgiving on the input data. The data can be comma or space separated. Extra spaces are no problem between data fields.

10/14/09

Changed configuration 0 sentence to output once per second even if input data stops coming in. TWC/LWC tab will indicate in yellow no data.

09/09/09

Reset the tab order for fields in Probe tab pane.

08/08/09

Changed received pressure from NOAA AOC data from "SP" to "PS" field.

07/21/09

Fixed data format parsing issue with NOAA AOC data.

114C Mansfield Hollow Road • P.O. Box 605
Mansfield Center, CT 06250-0605
Phone: (860) 450-1717 • Fax: (860) 450-1707
www.scieng.com

06/26/09

Removed the ProbeOn and DeiceOn parameters from the Probe eprom. We used to keep track of the time up to the minute for ProbeOn and DeiceOn elements. We had to remove this since the eproms don't support the necessary number of write cycles for this feature. This only applied to the Probe eprom, since it was the only one written to on a regular basis.

06/25/09

The zero was only done when in the TWC/LWC tab was in the front. This was because the power computation only got done when necessary to display the data in this tab. Modified SeaWcm to zero no matter which tab the user has in front.

06/24/09

Modified SeaWcm input data to allow for zero parameter. Both old and new formats are supported.

06/23/09

Fix issue with 'netsend' and 'netrecv' so it would not leave these running after they were no longer being used.

06/22/09

SeaWcm check for input data every half a second as opposed to once per second. Checking the data once per second is not sufficient and it causes the SeaWcm to get behind in case the input data is sent slightly faster than once per second. SeaWcm now reads to flush data from input port every time the user clicks on play. This is to prevent buffered data to get SeaWcm behind on input from port.

06/18/09

We had a client which required serial data to/from data system. They were not able to support network data connection to the WCM-2000. We made a major revision to the software to support this. Introduction of 'wcmread' to read data from socket/serial port. The 'wcmwrite' will do the write data to socket/serial port. The 'wcmrecord' will record data to a file. This last feature will be fully completed at a latter time. Modified SeaWcm interface to allow serial devices listed for input and output data ports. Also added file menu item to create and close a data file to write to.